CITREX[™]



«The compact and mobile testing device for ventilators.»

imtmedical

Developed for mobile use simple. compact. reliable.





Simple operation CITREX is simple and intuitive to operate. The colour screen offers excellent readability and can be adapted to any situation due to its flip-screen function.



Bidirectional flow measurement The newly developed measuring method allows extremely precise, bidirectional flow measurement with low measuring resistance.



Respiratory parameters All the relevant respiratory parameters are measured and calculated.



Real-time graphs The devices measures flow, volume, four pressures, temperature and oxygen concentration.





Gas standards and gas types 13 gas standards and 7 gas types can be measured so as to meet a range of measuring requirements.



Memory function It is simple to save measurements on the device and export them to external data media for subsequent analysis.



Interfaces Due to the numerous interfaces, the device is ideal for networking, remote control and configuration.



A compact device with everything you need CITREX is especially impressive due to its size, low weight and robustness. All required components are integrated and the battery enables prolonged independent use.

Options and accessories

The device is shipped with all important parts for immediate use in the field. There are also a number of accessories and options which can be purchased separately.



Adapter-Set (optional)

The adapters contained in the set allow connection of virtually any test object to the CITREX device. Minimum dead space and very slight differences in the diameter of the flow stream help increase measurement accuracy.

Price includes:

- CITREX H4 device
- High-performance battery
- Universal power plug
- USB cable
- Micro SD memory card
- Protection filter
- Quick-start guideData CD

Transport bag (optional)

The transport bag is made of high-quality materials and is big enough to securely hold and transport the device along with all accessories. There is also space for optional accessories such as test lungs and the adapter set.



Oxygen measurement (optional)

Fast and precise measurement of oxygen concentration is an important function when verifying and calibrating ventilators. This option is available for new devices or can be acquired subsequently as a retrofit set.



Technical Specifications


Flow and Pressur	re Measurements	Range	Accuracy
Flow		± 300 L/min	$\pm 1.9\%$ or $\pm 0.1 L/min$ **
Temperature compensated		yes	
Pressure compensated		yes	
Pressure			
High		010 bar	± 1%* or ± 10mbar**
Differential		± 200 mbar	± 0.75%* or ± 0.1 mbar**
Flow channel		-50150 mbar	± 0.75 %* or ± 0.1 mbar**
Barometer		5001150 mbar	± 1%* or ± 5 mbar**
Units			
Flow		L/min, L/s, cfm, mL/min, mL/s	
Pressure		bar, mbar, cmH ₂ O, inH ₂ O, Torr,	
		inHg, hPa, kPa, mmHg, PSI	
Other Measureme	ents	Range	Accuracy
Oxygen, pressure	compensated	0100%	± 1% O ₂ **
Gas temperature		050°C	± 1.75 %* or ± 0.5 °C**
Gas types		Air, Air/O ₂ , N ₂ O/O ₂ , Heliox	
		(21 % O ₂), He/O ₂ , N ₂ , CO ₂	
Gas standards		ATP, ATPD, ATPS, AP21, STP,	
		STPH, BTPS, BTPD, 0/1013, 20/981,	
		15/1013, 25/991, 20/1013	
Ventilation Parameters		Pongo	Accuracy
Ventilation Param	neters	Range	Accuracy
Breath rate	neters	11000 bpm	±1bpm or ± 2.5%**
	Ti,Te	11000 bpm 0.0560 s	±1bpm or ± 2.5%** ± 0.02s
Breath rate		11000 bpm	±1bpm or ± 2.5%** ± 0.02s ± 2.5%*
Breath rate Time	Ti,Te	11000 bpm 0.0560 s	±1bpm or ± 2.5%** ± 0.02s
Breath rate Time	Ti,Te I:E	11000 bpm 0.0560 s 1:300300:1	±1bpm or ± 2.5%** ± 0.02s ± 2.5%* ± 5%* ± 2%* or ± 20mL**
Breath rate Time Ratio	Ti,Te I:E Ti/Ttotal	11000 bpm 0.0560 s 1:300300:1 0100 %	±1bpm or ± 2.5%** ± 0.02s ± 2.5%* ± 5%*
Breath rate Time Ratio Volume	Ti,Te I:E Ti/Ttotal VTī, Vte Vi, Ve Insp. / Exp.	11000 bpm 0.0560 s 1:300300:1 0100 % ± 10 L	±1bpm or ± 2.5%** ± 0.02s ± 2.5%* ± 5%* ± 2%* or ± 20mL** ± 2.5%* ± 1.9%* or ± 0.1 L/min**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure	Ti,Te I:E Ti/Ttotal VTi, Vte Vi, Ve Insp. / Exp. Ppeak, Pmean, PEEP, Pplateau	11000 bpm 0.0560 s 1:300300:1 0100 % ± 10 L 0300 L/min	±1bpm or ± 2.5 %** ± 0.02 s ± 2.5 %* ± 5 %* ± 2 %* or ± 20 mL** ± 2.5 %* ± 1.9 %* or ± 0.1 L/min** ± 0.75 %* or ± 0.1 mbar**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure Compliance	Ti,Te I:E Ti/Ttotal VTī, Vte Vi, Ve Insp. / Exp.	11000 bpm 0.0560 s 1:300300:1 0100 % ± 10 L 0300 L/min ± 300 L/min	±1bpm or ± 2.5%** ± 0.02s ± 2.5%* ± 5%* ± 2%* or ± 20mL** ± 2.5%* ± 1.9%* or ± 0.1 L/min**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure	Ti,Te I:E Ti/Ttotal VTi, Vte Vi, Ve Insp. / Exp. Ppeak, Pmean, PEEP, Pplateau	11000 bpm 0.0560 s 1:300300:1 0100% ± 10L 0300 L/min ± 300 L/min 0150 mbar	±1bpm or ± 2.5 %** ± 0.02 s ± 2.5 %* ± 5 %* ± 2 %* or ± 20 mL** ± 2.5 %* ± 1.9 %* or ± 0.1 L/min** ± 0.75 %* or ± 0.1 mbar**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure Compliance	Ti,Te I:E Ti/Ttotal VTi, Vte Vi, Ve Insp. / Exp. Ppeak, Pmean, PEEP, Pplateau Cstat	11000 bpm 0.0560 s 1:300300:1 0100% ± 10L 0300 L/min ± 300 L/min 0150 mbar 01000 mL/mbar	±1bpm or ± 2.5 %** ± 0.02 s ± 2.5 %* ± 5 %* ± 2 %* or ± 20 mL** ± 2.5 %* ± 1.9 %* or ± 0.1 L/min** ± 0.75 %* or ± 0.1 mbar**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure Compliance	Ti,Te I:E Ti/Ttotal VTi, Vte Vi, Ve Insp. / Exp. Ppeak, Pmean, PEEP, Pplateau Cstat Adult, Pediatric, HFO	11000 bpm 0.0560 s 1:300300:1 0100% ± 10L 0300 L/min ± 300 L/min 0150 mbar 01000 mL/mbar flow or pressure at preset and	±1bpm or ± 2.5 %** ± 0.02 s ± 2.5 %* ± 5 %* ± 2 %* or ± 20 mL** ± 2.5 %* ± 1.9 %* or ± 0.1 L/min** ± 0.75 %* or ± 0.1 mbar**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure Compliance Volume trigger	Ti,Te I:E Ti/Ttotal VTi, Vte Vi, Ve Insp. / Exp. Ppeak, Pmean, PEEP, Pplateau Cstat Adult, Pediatric, HFO	11000 bpm 0.0560 s 1:300300:1 0100% ± 10L 0300 L/min ± 300 L/min 0150 mbar 01000 mL/mbar flow or pressure at preset and	±1bpm or ± 2.5 %** ± 0.02 s ± 2.5 %* ± 5 %* ± 2 %* or ± 20 mL** ± 2.5 %* ± 1.9 %* or ± 0.1 L/min** ± 0.75 %* or ± 0.1 mbar**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure Compliance Volume trigger General Informati	Ti,Te I:E Ti/Ttotal VTi, Vte Vi, Ve Insp. / Exp. Ppeak, Pmean, PEEP, Pplateau Cstat Adult, Pediatric, HFO	11000 bpm 0.0560 s 1:300300:1 0100 % ± 10L 0300L/min ± 300L/min 0150 mbar 01000 mL/mbar flow or pressure at preset and at adjustable levels yes flow, pressure, volume	±1bpm or ± 2.5 %** ± 0.02 s ± 2.5 %* ± 5 %* ± 2 %* or ± 20 mL** ± 2.5 %* ± 1.9 %* or ± 0.1 L/min** ± 0.75 %* or ± 0.1 mbar**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure Compliance Volume trigger General Informati Color display	Ti,Te I:E Ti/Ttotal VTi, Vte Vi, Ve Insp. / Exp. Ppeak, Pmean, PEEP, Pplateau Cstat Adult, Pediatric, HFO	11000 bpm 0.0560 s 1:300300:1 0100 % ± 10L 0300L/min ± 300L/min 0150 mbar 01000 mL/mbar flow or pressure at preset and at adjustable levels yes flow, pressure, volume RS-232, USB, Ethernet, CAN,	±1bpm or ± 2.5 %** ± 0.02 s ± 2.5 %* ± 5 %* ± 2 %* or ± 20 mL** ± 2.5 %* ± 1.9 %* or ± 0.1 L/min** ± 0.75 %* or ± 0.1 mbar**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure Compliance Volume trigger General Informati Color display Realtime curves Interface	Ti,Te I:E Ti/Ttotal VTi, Vte Vi, Ve Insp. / Exp. Ppeak, Pmean, PEEP, Pplateau Cstat Adult, Pediatric, HFO	11000 bpm 0.0560 s 1:300300:1 0100 % ± 10L 0300L/min ± 300L/min 0150 mbar 0150 mbar 01000 mL/mbar flow or pressure at preset and at adjustable levels yes flow, pressure, volume RS-232, USB, Ethernet, CAN, Analog Out, TTL	±1bpm or ± 2.5 %** ± 0.02 s ± 2.5 %* ± 5 %* ± 2 %* or ± 20 mL** ± 2.5 %* ± 1.9 %* or ± 0.1 L/min** ± 0.75 %* or ± 0.1 mbar**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure Compliance Volume trigger General Informati Color display Realtime curves Interface Power	Ti,Te I:E Ti/Ttotal VTi, Vte Vi, Ve Insp. / Exp. Ppeak, Pmean, PEEP, Pplateau Cstat Adult, Pediatric, HFO	11000 bpm 0.0560 s 1:300300:1 0100 % ± 10 L 0300 L/min ± 300 L/min 0150 mbar 0150 mbar 01000 mL/mbar flow or pressure at preset and at adjustable levels yes flow, pressure, volume RS-232, USB, Ethernet, CAN, Analog Out, TTL 90260 VAC	±1bpm or ± 2.5 %** ± 0.02 s ± 2.5 %* ± 5 %* ± 2 %* or ± 20 mL** ± 2.5 %* ± 1.9 %* or ± 0.1 L/min** ± 0.75 %* or ± 0.1 mbar**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure Compliance Volume trigger General Informati Color display Realtime curves Interface Power Battery	Ti,Te I:E Ti/Ttotal VTi, Vte Vi, Ve Insp. / Exp. Ppeak, Pmean, PEEP, Pplateau Cstat Adult, Pediatric, HFO	11000 bpm 0.0560 s 1:300300:1 0100 % ± 10 L 0300 L/min ± 300 L/min 0150 mbar 0150 mbar 01000 mL/mbar flow or pressure at preset and at adjustable levels yes flow, pressure, volume RS-232, USB, Ethernet, CAN, Analog Out, TTL 90260 VAC 4 hours	±1bpm or ± 2.5 %** ± 0.02 s ± 2.5 %* ± 5 %* ± 2 %* or ± 20 mL** ± 2.5 %* ± 1.9 %* or ± 0.1 L/min** ± 0.75 %* or ± 0.1 mbar**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure Compliance Volume trigger General Informati Color display Realtime curves Interface Power Battery Dimension (w×d×t	Ti,Te I:E Ti/Ttotal VTi, Vte Vi, Ve Insp. / Exp. Ppeak, Pmean, PEEP, Pplateau Cstat Adult, Pediatric, HFO	11000 bpm 0.0560 s 1:300300:1 0100 % ± 10 L 0300 L/min ± 300 L/min 0150 mbar 0150 mbar 01000 mL/mbar flow or pressure at preset and at adjustable levels yes flow, pressure, volume RS-232, USB, Ethernet, CAN, Analog Out, TTL 90260 VAC 4 hours 11.4 × 6 × 7 cm	±1bpm or ± 2.5 %** ± 0.02 s ± 2.5 %* ± 5 %* ± 2 %* or ± 20 mL** ± 2.5 %* ± 1.9 %* or ± 0.1 L/min** ± 0.75 %* or ± 0.1 mbar**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure Compliance Volume trigger General Informati Color display Realtime curves Interface Power Battery Dimension (w×d×h Weight	Ti,Te I:E Ti/Ttotal VTi, Vte Vi, Ve Insp. / Exp. Ppeak, Pmean, PEEP, Pplateau Cstat Adult, Pediatric, HFO	11000 bpm 0.0560 s 1:300300:1 0100% ± 10L 0300 L/min ± 300 L/min 0150 mbar 0150 mbar 01000 mL/mbar flow or pressure at preset and at adjustable levels yes flow, pressure, volume RS-232, USB, Ethernet, CAN, Analog Out, TTL 90260 VAC 4 hours 11.4 × 6 × 7 cm 0.4 kg	±1bpm or ± 2.5 %** ± 0.02 s ± 2.5 %* ± 5 %* ± 2 %* or ± 20 mL** ± 2.5 %* ± 1.9 %* or ± 0.1 L/min** ± 0.75 %* or ± 0.1 mbar**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure Compliance Volume trigger General Informati Color display Realtime curves Interface Power Battery Dimension (w×d×h Weight Calibration	Ti,Te I:E Ti/Ttotal VTi, Vte Vi, Ve Insp. / Exp. Ppeak, Pmean, PEEP, Pplateau Cstat Adult, Pediatric, HFO	11000 bpm 0.0560 s 1:300300:1 0100% ± 10L 0300 L/min ± 300 L/min 0150 mbar 01000 mL/mbar flow or pressure at preset and at adjustable levels yes flow, pressure, volume RS-232, USB, Ethernet, CAN, Analog Out, TTL 90260 VAC 4 hours 11.4 × 6 × 7 cm 0.4 kg Annually	±1bpm or ± 2.5 %** ± 0.02 s ± 2.5 %* ± 5 %* ± 2 %* or ± 20 mL** ± 2.5 %* ± 1.9 %* or ± 0.1 L/min** ± 0.75 %* or ± 0.1 mbar**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure Compliance Volume trigger General Informati Color display Realtime curves Interface Power Battery Dimension (w×d×h Weight Calibration Memory card	Ti,Te I:E Ti/Ttotal VTi, Vte Vi, Ve Insp. / Exp. Ppeak, Pmean, PEEP, Pplateau Cstat Adult, Pediatric, HFO	11000 bpm 0.0560 s 1:300300:1 0100% ± 10L 0300 L/min ± 300 L/min 0150 mbar 01000 mL/mbar flow or pressure at preset and at adjustable levels yes flow, pressure, volume RS-232, USB, Ethernet, CAN, Analog Out, TTL 90260 VAC 4 hours 11.4 × 6 × 7 cm 0.4 kg Annually yes	±1bpm or ± 2.5 %** ± 0.02 s ± 2.5 %* ± 5 %* ± 2 %* or ± 20 mL** ± 2.5 %* ± 1.9 %* or ± 0.1 L/min** ± 0.75 %* or ± 0.1 mbar**
Breath rate Time Ratio Volume Minute volume Peak flow Pressure Compliance Volume trigger General Informati Color display Realtime curves Interface Power Battery Dimension (w×d×h Weight Calibration	Ti,Te I:E Ti/Ttotal VTi, Vte Vi, Ve Insp. / Exp. Ppeak, Pmean, PEEP, Pplateau Cstat Adult, Pediatric, HFO	11000 bpm 0.0560 s 1:300300:1 0100% ± 10L 0300 L/min ± 300 L/min 0150 mbar 01000 mL/mbar flow or pressure at preset and at adjustable levels yes flow, pressure, volume RS-232, USB, Ethernet, CAN, Analog Out, TTL 90260 VAC 4 hours 11.4 × 6 × 7 cm 0.4 kg Annually	±1bpm or ± 2.5 %** ± 0.02 s ± 2.5 %* ± 5 %* ± 2 %* or ± 20 mL** ± 2.5 %* ± 1.9 %* or ± 0.1 L/min** ± 0.75 %* or ± 0.1 mbar**

The perfect device for every application

For a several years, now, imtmedical has been the market leader and most important supplier of testing and calibration solutions for ventilators and anaesthesia devices. Developers appreciate the reliability and accuracy of the devices, as do service technicians and quality specialists.







CITREX was designed for mobile use and meets all the requirements of day-to-day field operations. Reliable, compact and mobile.

- Verification and calibration of ventilators (hospital and homecare)
- Use in production plants

The market leader in the lab and development category

The three models of the FlowAnalyser are put to use wherever high-precision measurement of pressure, flow and volume is required. Measurements can be subjected to detailed analysis using FlowLab software.

- · Lab, research and development
- Calibration of ventilators
- Anaesthesia gas measurement
- Verification of spirometers and oxygen concentrators



...and just the right test lung, too

The various imtmedical test lung models cover every conceivable purpose.

- · Calibration of ventilators and anaesthetic equipment
- Quick daily check of various devices
- Instruction courses and training programs
- Research and development

